

Report of Calibration

Taisho Biomed Instruments Co., Ltd.
1-11-19 Sagisu, Fukushima-ku
Osaka 553-0002 Japan



UNIT UNDER TEST:	Fluke Biomedical ESA612 Electrical Safety Analyzer	TEST_RESULT:	Pass
SERIAL NUMBER:		PERFORMED ON:	
ASSET NUMBER:		DATA TYPE:	Found Left
PROCEDURE NAME:	ESA612B VER2 - Exp Unc WW	TEMPERATURE:	25.8 °C
PROCEDURE_REV:	6	HUMIDITY:	70.3 %
CALIBRATED BY:			
PO NUMBER:			

Taisho Biomed Instruments Co., Ltd. certifies that the above listed instrument meets or exceeds all specifications as stated in the referenced procedure unless otherwise noted. It has been calibrated using measurement standards traceable to the National Institute of Standards and Technology (NIST), or to NIST accepted intrinsic standards of measurement, or derived by the ratio type of self-calibration techniques. This calibration complies with MIL-STD-45662A and ANSI/NCSL Z540.1-1999 (R2002).

This report may not be reproduced, except in full, unless permission for the publication of an approved abstract is obtained in writing from the calibration organization issuing this report.

Note: Any Test Uncertainty Ratio (TUR) that is less than four to one will appear under the "TUR" heading on the data record. If the TUR meets or exceeds four to one, the field is left blank.

REMARKS:

Standards Used

<u>Asset #</u>	<u>Description</u>	<u>Cal Date</u>	<u>Due Date</u>
001		8/7/2024	8/7/2025
006		9/5/2024	9/5/2025
010		7/22/2024	7/22/2025
014		7/22/2024	7/22/2025
015		7/16/2024	7/16/2025
016		8/26/2024	8/26/2025
017		8/9/2024	8/9/2025
020		7/18/2024	7/18/2025
PRE0000923		8/13/2024	8/13/2025
PRE0000924		7/10/2024	7/10/2025

Test Results

<u>Test Description</u>	<u>True Value</u>	<u>Test Results</u>	<u>Lower Limit</u>	<u>Upper Limit</u>	<u>TUR</u>
OUTLET TENSION TESTS					
Equipment Outlet L1 slot tension is greater than 16 oz (4.45 N)					Pass
Equipment Outlet L2 slot tension is greater than 16 oz (4.45 N)					Pass
Equipment Outlet Ground tension is greater than 8 oz (2.22 N)					Pass
UUT is ESA612B, UI-3.01.03, MTR-3.01.03, VNC-1.01.52, SN-As-Found NOMINAL: 100 Measured line voltage = 100.06 V UUT Current Draw: 0.14 A					
FUNCTIONAL TESTS					
Warning LED ON/OFF Test					Pass
UUT Beeper ON/OFF Test					Pass
UUT Keypad Tests					Pass
EQUIPMENT OUTLET POLARITY TESTS					
Normal Polarity, Closed Neutral, Closed Earth					Pass
Normal Polarity, Closed Neutral, Open Earth					Pass

大正医科器械株式会社 大阪福島セントラルビル
大阪市福島区鷺洲 1-11-19

TEL. 06-6451-7177

FAX. 06-6451-7178

Internet <https://taishobiomed.com>

Print Date:

Test Results

Test Description	True Value	Test Results	Lower Limit	Upper Limit	TUR
Normal Polarity, Open Neutral, Closed Earth					Pass
Reversed Polarity, Closed Neutral, Closed Earth					Pass
Reversed Polarity, Open Neutral, Closed Earth					Pass
Reversed Polarity, Closed Neutral, Open Earth					Pass
EQUIPMENT OUTLET GROUND TO NULL JACK RESISTANCE TEST					
EO Ground to NULL Jack Open (>1 GΩ)					Pass
0.000 Ω	0.000 Ω	0.5041 Ω	-2.5000 Ω	2.5000 Ω	Pass
MAINS VOLTAGE ACCURACY					
Measure L1-L2	100.04 V	99.9 V	97.8 V	102.2 V	Pass
Measure L1-GND	100.06 V	99.9 V	97.9 V	102.3 V	Pass
GFI VERIFICATION					
5 mA Trip Point					Pass
4.5 mA Test (No Trip)					Pass
5.5 mA Test (Tripped)					Pass
GFI Error Test					Pass
10 mA Trip Point					Pass
9 mA Test (No Trip)					Pass
11 mA Test (Tripped)					Pass
25 mA Trip Point					Pass
22.5 mA Test (No Trip)					Pass
27.5 mA Test (Tripped)					Pass
POINT-TO-POINT VOLTAGE VERIFICATION					
1.000 V	1.0000 V	1.000 V	0.780 V	1.220 V	Pass
8.000 V	8.0000 V	8.005 V	7.640 V	8.360 V	Pass
25.00 V	25.000 V	25.01 V	24.30 V	25.70 V	Pass
80.00 V	80.000 V	80.02 V	78.20 V	81.80 V	Pass
240.0 V	240.00 V	240.1 V	235.0 V	245.0 V	Pass
POINT-TO-POINT RESISTANCE VERIFICATION					
Red/Black Jack Input					
0.000 Ω	0.0000 Ω	0.000 Ω	-0.015 Ω	0.015 Ω	Pass
0.200 Ω	0.2000 Ω	0.198 Ω	0.181 Ω	0.219 Ω	Pass
1.800 Ω	1.8000 Ω	1.796 Ω	1.749 Ω	1.851 Ω	Pass
Red Jack/EO Ground Terminal					
0.000 Ω	0.0000 Ω	0.001 Ω	-0.015 Ω	0.015 Ω	Pass
1.800 Ω	1.8000 Ω	1.802 Ω	1.749 Ω	1.851 Ω	Pass
INSULATION RESISTANCE VERIFICATION					
250 V Source	250.0 V	262.67 V	250.00 V	300.00 V	Pass
500 V Source	500.0 V	524.28 V	500.00 V	600.00 V	Pass
2 mA Current Limit	2.000 mA	1.9352 mA	1.7500 mA	2.2500 mA	Pass
Mains-PE (500V)	10.0000 MΩ	9.800 MΩ	9.600 MΩ	10.400 MΩ	Pass
AP-PE (500V)	10.0000 MΩ	9.992 MΩ	9.600 MΩ	10.400 MΩ	Pass
AP-NE (500V)	10.0000 MΩ	9.995 MΩ	9.600 MΩ	10.400 MΩ	Pass
Mains-NE (500V)	10.0000 MΩ	9.994 MΩ	9.600 MΩ	10.400 MΩ	Pass
Mains-AP (RA) (500V)	10.0000 MΩ	9.995 MΩ	9.600 MΩ	10.400 MΩ	Pass
Mains-AP (LL) (500V)	10.0000 MΩ	9.995 MΩ	9.600 MΩ	10.400 MΩ	Pass
Mains-AP (LA) (500V)	10.0000 MΩ	9.992 MΩ	9.600 MΩ	10.400 MΩ	Pass
Mains-AP (RL) (500V)	10.0000 MΩ	9.992 MΩ	9.600 MΩ	10.400 MΩ	Pass
Mains-AP (V1) (500V)	10.0000 MΩ	9.992 MΩ	9.600 MΩ	10.400 MΩ	Pass
INSULATION RESISTANCE RANGE VERIFICATION					
Mains-AP (V1) (250V)	0.70000 MΩ	0.7003 MΩ	0.4860 MΩ	0.9140 MΩ	Pass
Mains-AP (V1) (500V)	1.00000 MΩ	0.9997 MΩ	0.7800 MΩ	1.2200 MΩ	Pass
Mains-AP (V1) (250V)	6.50000 MΩ	6.4950 MΩ	6.1700 MΩ	6.8300 MΩ	Pass
Mains-AP (V1) (500V)	18.0000 MΩ	17.987 MΩ	17.440 MΩ	18.560 MΩ	Pass

Test Results

<u>Test Description</u>	<u>True Value</u>	<u>Test Results</u>	<u>Lower Limit</u>	<u>Upper Limit</u>	<u>TUR</u>
Mains-AP (V1) (250V)	22.0000 MΩ	22.004 MΩ	20.150 MΩ	23.850 MΩ	Pass
Mains-AP (V1) (500V)	60.0000 MΩ	59.923 MΩ	55.300 MΩ	64.700 MΩ	Pass
Mains-AP (V1) (250V)	100.000 MΩ	100.19 MΩ	92.30 MΩ	107.70 MΩ	Pass

DC LEAKAGE VERIFICATION

AAMI Load Resistance Test

1.000 kΩ	1.000 kΩ	1.0007 kΩ	0.9800 kΩ	1.0200 kΩ	Pass
----------	----------	-----------	-----------	-----------	------

DC Leakage Current Tests

10.0 uA	10.00 μA	9.9 μA	8.9 μA	11.1 μA	Pass
50.0 uA	50.00 μA	49.9 μA	48.5 μA	51.5 μA	Pass
100.0 uA	100.00 μA	99.9 μA	98.0 μA	102.0 μA	Pass
500 uA	500.0 μA	500 μA	494 μA	506 μA	Pass
1600 uA	1600.0 μA	1600 μA	1583 μA	1617 μA	Pass
7.000 mA	7.000 mA	7.00 mA	6.92 mA	7.08 mA	Pass

AC LEAKAGE FILTER FREQUENCY RESPONSE (AAMI Load)

1000 uA, 60 Hz	998.00 μA	999.7 μA	987.0 μA	1009.0 μA	Pass	
1000 uA, 1 kHz	691.00 μA	688.0 μA	683.1 μA	698.9 μA	Pass	
1000 uA, 10 kHz	95.60 μA	94.7 μA	92.7 μA	98.5 μA	Pass	2.19
1000 uA, 30 kHz	33.10 μA	32.8 μA	31.4 μA	34.8 μA	Pass	2.32

DIFFERENTIAL LEAKAGE VERIFICATION - NORMAL POLARITY

(75 to 199)uA Range

76 uA @ 60 Hz	76.1 μA	76 μA	48 μA	104 μA	Pass
160 uA @ 60 Hz	160.2 μA	162 μA	124 μA	196 μA	Pass

(200 to 1999)uA Range

500 uA @ 60 Hz	499.4 μA	509 μA	429 μA	569 μA	Pass
1600 uA @ 60 Hz	1600.3 μA	1634 μA	1420 μA	1780 μA	Pass

(2 to 20)mA Range

5.00 mA @ 60 Hz	5.007 mA	5.07 mA	4.49 mA	5.53 mA	Pass
16.00 mA @ 60 Hz	15.995 mA	16.04 mA	14.38 mA	17.61 mA	Pass

DIFFERENTIAL LEAKAGE VERIFICATION - REVERSE POLARITY

(2 to 20)mA Range

16.00 mA @ 60 Hz	15.997 mA	16.05 mA	14.38 mA	17.62 mA	Pass
------------------	-----------	----------	----------	----------	------

ECG LEAKAGE FUNCTIONALITY VERIFICATION

Applied Part (RA), 101 uA Detected					Pass
Applied Part (RA) Open Measures > 1 GigaΩ					Pass
Applied Part (LL), 101 uA Detected					Pass
Applied Part (LL) Open Measures > 1 GigaΩ					Pass
Applied Part (LA), 101 uA Detected					Pass
Applied Part (LA) Open Measures > 1 GigaΩ					Pass
Applied Part (RL), 101 uA Detected					Pass
Applied Part (RL) Open Measures > 1 GigaΩ					Pass
Applied Part (V1), 101 uA Detected					Pass
Applied Part (V1) Open Measures > 1 GigaΩ					Pass

DIRECT APPLIED PART LEAKAGE - STD 353 VERIFICATION

GND to RA

1.000 mA @ 60 Hz	0.8637 mA	0.861 mA	0.854 mA	0.873 mA	Pass
------------------	-----------	----------	----------	----------	------

RED to RA

1.000 mA @ 60 Hz	0.8629 mA	0.860 mA	0.853 mA	0.873 mA	Pass
------------------	-----------	----------	----------	----------	------

ALTERNATIVE EQUIPMENT LEAKAGE

RED to HOT

1.000 mA @ 60 Hz	0.8685 mA	0.868 mA	0.859 mA	0.878 mA	Pass
------------------	-----------	----------	----------	----------	------

ALTERNATIVE EQUIPMENT, APPLIED PART LEAKAGE

RA to HOT (SAF)

Test Results

<u>Test Description</u>	<u>True Value</u>	<u>Test Results</u>	<u>Lower Limit</u>	<u>Upper Limit</u>	<u>TUR</u>
1.000 mA @ 60 Hz	0.8697 mA	0.868 mA	0.860 mA	0.879 mA	Pass
ALTERNATIVE APPLIED PART PATIENT LEAKAGE					
RA to NEUTRAL (SPAT)					
1.000 mA @ 60 Hz	0.8666 mA	0.864 mA	0.857 mA	0.876 mA	Pass
PATIENT AUXILIARY LEAKAGE - STD 601					
RA to RL					
0.1000 mA @ 100 Hz	0.10004 mA	0.0997 mA	0.0980 mA	0.1020 mA	Pass
DIRECT EQUIPMENT, PATIENT LEAKAGE - STD 353					
GND to RA					
0.1000 mA @ 100 Hz	0.10004 mA	0.0997 mA	0.0980 mA	0.1020 mA	Pass
DIRECT EQUIPMENT, PATIENT LEAKAGE - STD 601					
GND to RA					
0.1000 mA @ 100 Hz	0.10004 mA	0.0997 mA	0.0980 mA	0.1020 mA	Pass
DIRECT EQUIPMENT, ENCLOSURE LEAKAGE - STD 353					
GND to RED					
0.1000 mA @ 100 Hz	0.10004 mA	0.0997 mA	0.0980 mA	0.1020 mA	Pass
DIRECT EQUIPMENT, ENCLOSURE LEAKAGE - STD 601					
GND to RED					
0.1000 mA @ 100 Hz	0.10004 mA	0.0997 mA	0.0980 mA	0.1020 mA	Pass
FILTER TESTS					
Apply 100Hz, Measure DC					
0.00 uA @ 100 Hz	0.000 uA	0.00 uA	0.00 uA	1.00 uA	Pass
Apply DC, Measure AC					
0.00 uA	0.000 uA	0.00 uA	0.00 uA	1.00 uA	Pass
MAP VOLTAGE AND CURRENT LIMIT TESTS					
120.0 V @ 60 Hz	120.10 V	120.0 V	111.7 V	128.5 V	Pass
230.0 V @ 50 Hz	230.10 V	230.1 V	214.0 V	246.2 V	Pass
3.50 mA @ 50 Hz	3.472 mA	3.47 mA	2.60 mA	4.34 mA	Pass
7.50 mA @ 50 Hz	7.183 mA	7.17 mA	5.39 mA	8.98 mA	Pass
1.00 mA @ 60 Hz	1.040 mA	1.03 mA	0.78 mA	1.30 mA	Pass
ECG WAVEFORM TESTS - 2 Hz SQUAREWAVE					
2.000 Hz	2.000 Hz	2.0000 Hz	1.9600 Hz	2.0400 Hz	Pass
RA-RL Amplitude	0.674 mV	0.6879 mV	0.6403 mV	0.7077 mV	Pass 2.29
LL-RL Amplitude	1.673 mV	1.7123 mV	1.5894 mV	1.7567 mV	Pass
LA-RL Amplitude	1.384 mV	1.4128 mV	1.3148 mV	1.4532 mV	Pass
V1-RL Amplitude	2.074 mV	2.1208 mV	1.9703 mV	2.1777 mV	Pass
EQUIPMENT CURRENT VERIFICATION					
Test 1	1.0060 A	1.000 A	0.756 A	1.256 A	Pass 3.38
Test 2	3.0094 A	3.000 A	2.659 A	3.360 A	Pass

Cal Info SN: CAL DATE:

*** End of Procedure ***

**** End of Certificate ****